



Injection 1 µl, split 1:20, column: ZB624, 60 m, 0.32 mm, 1.8 µm, 120 kpa Helium
 Temp. program: 40 °C / 6 min > 20 °/min > 260 °C / 13 min (30 min)
 EPED: 100 µm, 4 Hz, 50 ms, O₂: 10 kpa, H₂: 10 kpa
 Analog out: Cl*3, multiplier 1

No.	Substances	(ng Cl)	No.	Substances	(ng Cl)
1	Trichlorotrifluoroethane	0,57	8	Tetrachloroethene	0,76
2	Methylene chlorine	0,60	9	Chlorobenzene	0,19
3	Chloroforme	0,72	10	1,2,3-Trichloropropane	0,55
4	1,1,1-Trichloroethane	0,58	11	2-Chlorotoluene	0,17
5	Carbon Tetrachlorine	0,80	12	4-Chlorotoluene	0,16
6	Bromodichloromethane	0,79	13	1,3-Dichlorobenzene	0,34
7	Dibromochloromethane	0,32	14	1,2-Dibromo-3-Chlorpropane	0,28